



The American Association for Laboratory Accreditation

World Class Accreditation

# Accredited Laboratory

A2LA has accredited

## THE L.S. STARRETT COMPANY, TRU-STONE TECHNOLOGIES DIVISION

*Waite Park, MN*

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets the requirements of additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009*).



Presented this 5th day of January 2010.

  
\_\_\_\_\_  
President & CEO

For the Accreditation Council  
Certificate Number 1580.01  
Valid to November 30, 2011

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

THE L.S. STARRETT COMPANY,  
 TRU-STONE TECHNOLOGIES DIVISION  
 1101 Prosper Drive  
 P.O. Box 430  
 Waite Park, MN 56387  
 Ken Barros Phone: 320 251 7171

CALIBRATION

Valid To: November 30, 2011

Certificate Number: 1580.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

I. Dimensional

Parameter/Equipment	Range	CMC <sup>2,3</sup> (±)	Comments
Granite Surface Plates –  Flatness	±25 µin over 12 in Up to 300 inch diagonal	(5.2√D) µin	Autocollimator
	0.020 in per measurement Up to 480 inch diagonal	(6.2√D) µin	Electronic level
Repeat Reading	Up to 0.002 in	25 µin	Repeat-o-meter with 0.00002 indicator  Only valid in connection with a flatness or straightness calibration

Parameter/Equipment	Range	CMC <sup>2,3</sup> (±)	Comments	
Granite Metrology Accessories –	Flatness/Straightness			
		Up to 0.0008 in	19 µin	Mikrokator and straight edge
		±25 µin over 12 in Up to 300 inch diagonal	(5.2√D) µin	Autocollimator
		0.020 in per measurement Up to 480 inch diagonal	(6.2√D) µin	Electronic level
	Perpendicularity	≤ 48 in	2.5 µin/in	Pentaprism and autocollimator
Parallelism		Up to 0.001 in	26 µin	Electronic amp and gage head
		0.020 in per measurement Up to 480 inch length	(6.2√D) µin	Electronic level

<sup>1</sup> This laboratory offers commercial calibration service.

<sup>2</sup> Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> In the statement of CMC,  $D$  is the length of the diagonal in inches.